



# California Medical Device Recall Information



## Recall Name

### Med-Vet International Recalls Veterinary 1/2cc U-40 Insulin Syringes Due to Mismatched Syringe Barrels

Recall Date	Product Description	Recalling Firm	Recall Reason
2/14/14	Veterinary 1/2cc U-40 Insulin Syringes  140 boxes; 29g x 1/2" needle	<b>Med-Vet International</b> Mettawa, IL	<i>Potential for syringes to be incorrectly labeled with "40 units per 1/2cc"; instead, they should be marked with only 20 units per 1/2cc.</i>
Recall Class	Product Identification	Distribution	Affected Dates
N/A	Suspected Lot recalled: <ul style="list-style-type: none"><li>• <b>20120610</b></li></ul> Item Number: <ul style="list-style-type: none"><li>• <b>"MV1/2CCINS-40"</b></li><li>• <b>"1/2CCINS-40 by Oasis"</b></li></ul> <a href="#">Product Labels</a>	<b>CA</b> , nationwide	Distributed from:  January 2013 to February 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm386613.htm>